


The Rotex Screw Needle Biopsy Instrument [®] CE 0413



The original Rotex-Nordenström instrument for sampling of cellular material was first introduced on the market by URSUS in 1975. It consists of a thin screw needle housed in a cannula with an outer diameter of either 0.8 or 1 mm and a special instrument holder. The advantages of the screw needle technique were soon recognized and were pointed out in comparative studies by House & Thomson(2), Sinner(7) and others. The rapid acceptance of the new technique in leading radiological and cytological quarters has thereafter led to further refinement of the instrument.

Construction of the instrument


The screw needle (fig.B) is made of a 0.55 mm thick stainless steel rod with a small handle at its proximal end. The distal 16 mm of the rod have been formed into a tapering

screw with cutting ridge (Fig. A). The needle is housed in a steel cannula with an outer diameter of 0.8 or 1.0 mm (equivalent to 21-22 gauge needles), respectively. The length of the cannula (Fig. C) matches that of the screw needle. Despite its small diameter, the screw needle with the 0.8 mm cannula is sufficiently stiff to be inserted for sampling of cellular material from tissue. The plastic hub of the cannula has the shape of a helical screw, which fits into the teeth of the accompanying instrument holder (Fig. E). The screw needle and cannula are held in place by the instrument holder's hinged lock (Fig. D). The instrument is inserted with the screw needle in a protected position in the cannula.

The screw needle is also positioned in the cannula at the withdrawal of the instrument. The risk of seeding surrounding tissues with cellular material from the screw needle during withdrawal of the instrument is thereby reduced to a minimum. The screw needle and cannula are delivered in standard packs of 20 instruments, each in individual protective packaging, sterilized and ready for use. They are disposable and intended for single use to ensure efficient sampling, sterility and indisputable security against any kind of accidental inoculation. The instrument holder, which is made of polycarbonate, is intended for repeated use and is sterilisable by heat (max. 120 degrees C, 15 minutes) or by gas. Each standard pack of 20 instruments includes two instrument holders. The instrument is available in other standard dimensions according to our price-list.

Assembly of the instrument

- 1) Open the hinged lock of the instrument holder.
- 2) Remove the screw needle together with its cannula housing from its protective packaging and place it in the two slots in the head of the instrument holder so that the cannula's helical hub assumes the position against the back wall of the instrument holder (Fig. D and F). Pull back the handle of the screw needle so that its screw is entirely withdrawn into the cannula.
- 3) Swing the hinged lock up to the locked position, thereby locking the cannula hub against the back wall of the instrument holder. The instrument is now ready for use.



Preparation of the insertion site

- 1) After selecting an appropriate insertion site, clean and anaesthetize the skin in the underlying tissue.
- 2) An ordinary disposable 1.5 mm thick needle is then used to make a puncture canal through the skin.



Sampling cell material

- 1) Introduce the cannula of the assembled Rotex screw needle biopsy instrument into the puncture canal and guide it up to the edge of the tissue to be sampled (Fig. F), under guidance of biplane fluoroscopy, CT or ultrasound.
- 2) Rotate the handle of the screw needle clockwise, thereby driving the 16 mm long screwed part into the tissue to be sampled (Fig. G). In the event strong resistance is encountered, rotate the screw needle no more than 5 turns. This is done in order to prevent the screw needle from bending. Pull

the needle back into the cannula, push it forward again, and rotate it another 5 turns into the tissue. Repeat this procedure until the full length of the screwed part of the needle has been introduced into the lesion. In lesions with soft material, the screw needle is pushed and pulled in and out of the tissue 2-3 times and then secured in its protected position in the cannula.

- 3) The helical hub screw of the cannula is, after the screwed part of the needle is driven into the tissue, rotated anti-clockwise until the screwed part is positioned in the protected position of the cannula (Fig. H). The cannula now holds the screw needle, which contains tissue and cell material from the sample in its grooves in a protected position. Increasing resistance to rotation of the helical hub may indicate that the screw needle tip is bent. Forcing rotation of the hub may then lead to further bending of the screw tip. When in doubt, the screw needle should simply be pulled back into the cannula.

- 4) When sampling has been completed, withdraw the instrument and release the cannula and screw needle from the instrument holder.



Retrieval of collected cell material

- 1) Push the screw needle out of the tip of the cannula. Place the screw against a glass slide

(Fig. I) and prepare smears from the material, which are then fixed and stained.

- 2) Material rolled up in the grooves of the screw is retrieved by rotating the screw needle anti-clockwise against the edge of a

sterile glass slide (Fig. J). The fragments are then fixed in formaline. Cell and tissue material thus obtained may be imbedded and treated as an ordinary histological specimen.

3) After removal of the screw needle from the cannula an ordinary syringe should be

Some advantages

- ◆ The Rotex Screw Needle Biopsy Instrument offers a safe, simple way of obtaining highly reliable samples. Since material is excised instead of being aspirated, complete cell clusters are obtained rather than single cells. As a result, at least 95% diagnostic accuracy has been obtained using the Rotex Screw Needle Biopsy Instrument for sampling lung tumors (2) and non palpable breast tumors (9). The diameter of the instrument cannula is 0.8 or 1.0 mm, which explains the low incidence of reported complications (2)
- ◆ The echogenic Rotex Screw Needle Biopsy Instrument improves sampling accuracy. The tip of the screw needle is clearly visible during sonographic guidance.
- ◆ The instrument is easy to operate. It occupies a minimum of space under an image intensifier.
- ◆ Sampling is performed efficiently over a distance of 16 mm, corresponding to the spiral portion of the screw needle. This enhances the possibility of obtaining representative samples. The screw needle and cannula work together as a cutting instrument, and the screw needle stabilize the tissue when the cannula is rotated over it. In this way all tissue components are sampled before they are lifted out. Material from fibrous inflammatory lesions and highly organized tissue such as benign tumors can therefore also be obtained. The material is highly concentrated and is not diluted by tissue fluid and blood as in aspiration biopsy. This facilitates microscopic examination of the material. Rather large flakes of cells and connective tissue can often be obtained. The clumps of tissue which can be sampled from organized tissue may allow treatment of the material with histological techniques.

attached to the cannula and its contents blown out. This material may then be smeared on a glass slide for staining and microscopy or transferred to a bacteriological culture medium.

- ◆ Material from the screw needle can be inoculated directly on culture media for bacteriological examination.
- ◆ The instrument holder allows the examiner to keep his hand out of the X-ray beam.
- ◆ The dimensions of the instrument have been optimized with regard to minimizing trauma and maximizing stability. This latter factor is of importance in order to enable precision sampling. In this way unnecessary insertion of the cannula in several directions can be avoided.
- ◆ The screw needle is always in a protected position; i.e., withdrawn into the cannula, when the helical hub is positioned distally in the instrument holder. In this way the risk of tumor seeding to surrounding tissue is reduced to a minimum.

Fields of application

The Rotex instrument was originally constructed for the biopsy of lung lesions. It has, however, proven to be equally suitable for the biopsy of other organs such as the liver, the kidney, lymph nodes, mammary tissue, etc. Even malignant bone lesions may be biopsied with the instrument. In this case a guide cannula 1.5 mm in diameter is first introduced through the overlying soft tissue and normal bone structures. The Rotex screw needle biopsy instrument is then inserted through this cannula. Sampling of material may then be performed as described above for soft tissue biopsies.

WARNING

The Rotex Screw Needle Biopsy Instrument is for single-use only and must not be reused. Reusing arouses risks of infection, contamination, tumor seeding and failure in usability and transfer all legal responsibility to the user.

References

1. Nordenström, B.: A new instrument for biopsy. *Radiology* 117 (1975), 474-475.
2. Sinner, W.N.: The diagnosis of pulmonary lesions by percutaneous transthoracic needle biopsy. Thesis, Karolinska institutet, Stockholm 1976.
3. House, A.J.S. and Thomson, K.R.: Evaluation of a new transthoracic needle for biopsy of benign and malignant lung lesions. *Am. J. Roentgenol.* 129 (1977), 215-220.
4. Nordenström, B. and Sinner, W.N.: Needle biopsies of pulmonary lesions. *Fortschr. Röntgenstr.* 129 (1978), 414-418.
5. Nordenström, B. and Sinner, W.N.: Early diagnosis of malignant pulmonary lesions. *Radiologie* 19 (1979), 162-168.
6. Nordenström, B.: Transthoracic needle biopsy. In: *Percutaneous biopsy and therapeutic vascular occlusion*. Eds. Anacker, H., Gullota, U., Rupp, N. Thieme Verlag, Stuttgart 1980, pp. 11-19.
7. Allison, D.J. and Hemingway, A.P.: Percutaneous needle biopsy of the Lung, *British Medical J.* Vol.282, 875.
8. Svane, G.: Stereotaxic needle biopsy of non-palpable breast lesions. *Acta Radiol. Diagn.* 1983, 24: 385-390.
9. Gent, H.J., Sprenger, E., Dowlatshahi, K.: Stereotaxic Needle Localization and Cytological Diagnosis of Occult Breast Lesions. *Ann. Surg.* 204, 1986, 580-584.
10. Diethelm, L., Heuck, F., Olsson, O., Strnad, H., Zuppinger, A.(eds.): *Encyclopedia of Medical Radiology*. Vol. IX, Part 5B; Nordenström, B.: Needle Biopsy of Pulmonary Lesion. Springer-Verlag, Berlin, Heidelberg 1989, pp 439-469.
11. Takahashi, M., Sano, A., Nishizawa, S., et al. CT-guided biopsy of thoracic mass lesions followed by fast stain technique. *Nippon Igaku Hoshasen Gakkai Zasshi.* 1990 Mar 25;50 (3): 249-57.
12. Kosnik, M., Suskovic, S. Comparison of the results of transthoracic needle biopsy of the lungs using the Nordenström and the Rotex needles. *Plucne Bolesti.* 1990 Jul-Dec;42 (3-4):174-7.
13. Sanders, W.H. and Lampmann, L.E. Percutaneous ultrasound guided management of pericardial fluid. *Eur J Radiol.* 1991 Mar-Apr;12 (2):147-9.
14. Szolar, D.H, Preidler, K.W., Kugler, C. et al. Fluoroscopically guided fine-needle pulmonary biopsy using the Rotex needle. *Fortschr. Röntgenstr.* 1994;161 (12). 505-511.
15. Munshi, M., Shrivastava, S., Agrawal, S.V., et al. Cytodiagnosis of lower respiratory tract lesions by transthoracic needle aspiration. *Indian J Pathol Microbiol.* 1995 Oct, 38 (4):417-21.
16. Welker, L., Akkan, R., Holz, O., Schultz, H., Magnussen, H. Diagnostic outcome of two different CT-guided fine needle biopsy procedures. *Diagnostic Pathology* 2007 Aug. 23;2:31.

The Rotex Screw Needle Biopsy Instrument® is manufactured by

URSUS
MEDICAL

*Ursus Medical AB, Arsenalsgatan 4, SE-111 47 Stockholm, Sweden
Telephone + 46 8 679 74 75, Telefax +46 8 679 72 20
E-mail: info@ursus.se
Web: www.ursus.se*